



Certificate of Analysis

Compliance Test

Client Information:

Honest PP&D

**1038 Arlington St
Orlando, Florida 32805**

Order # HON250410-010004

Order Date: 2025-04-10

Sample # AAGP159

Batch # GYDHJWM252401

Batch Date: 2025-03-26

Extracted From: HEMP

Test Reg State: Georgia

Sampling Date: 2025-04-11

Lab Batch Date: 2025-04-11

Completion Date: 2025-04-16

Initial Gross Weight: 60.400 g



Product Image



Pathogenic Microbiology

Tested



Heavy Metals

Passed



Mycotoxins

Passed



Pesticides

Passed



Residual Solvents

Passed



Microbiology

Tested



Filtration and Foreign

Passed



Microbiology Pathogenic AE qPCR - (GA)

Specimen Weight: 1038.300 mg

Dilution Factor: 1.000

Analyte	Result (cfu/g)	Analyte	Result (cfu/g)
Aspergillus (Flavus, Fumigatus, Niger, Terreus)	Passed	STEC	Passed

Tested
SOP13.029 (qPCR)



Microbiology AC TYM BTGN plating - (GA)

Specimen Weight: 993.900 mg

Dilution Factor: 1.000

Analyte	LOQ (cfu/g)	Result (cfu/g)	Analyte	LOQ (cfu/g)	Result (cfu/g)
Bile tolerant gram-negative bacteria	100	<100	Total Yeast/Mold	1000	<1000
Total Aerobic Count	1000	1000.0			

Tested
SOP13.003 (Petrifilm)



Filtration and Foreign Material - (GA)

Specimen Weight: **N/A** Dilution Factor: 1.000

Passed
SOP13.020 (Electronic Balance)

Analyte	Action Level (%)	Result (%)	Analyte	Action Level (%)	Result (%)
Covered Area	10	0.000	Weight %	1	0.000
Feces	0.5	0.000			


Aixa Sun Lab Director/Principal Scientist

D.H.Sc., M.Sc., B.Sc., MT (AAB)



Definitions and Abbreviations used in this report: Total Active CBD = CBD + (CBD-A * 0.877), *Total CBDV = CBDV + (CBDVA * 0.867), Total Active THC = THCA-A * 0.877 + Delta 9 THC, Total THCV = THCV + (THCVA * 0.87), CBG Total = (CBGA * 0.878) + CBG, CBN Total = (CBNA * 0.876) + CBN, Total CBC = CBC + (CBCA * 0.877), Total THC-O-Acetate = Delta 8 THC-O-Acetate + Delta 9 THC-O-Acetate, Total THCP = Delta8-THCP + Delta9-THCP, Total Cannabinoids = Total percentage of cannabinoids within the sample. (mg/ml) = Milligrams per Milliliter, LOQ = Limit of Quantitation, LOD = Limit of Detection, Dilution = Dilution Factor, (ppb) = Parts per Billion, (%) = Percent, (cfu/g) = Colony Forming Unit per Gram, (µg/g) = Microgram per Gram, (ppm) = Parts per Million, (ppm) = (µg/g), (aw) = Water Activity, (mg/Kg) = Milligram per Kilogram, The results apply to the sample as received.
This report shall not be reproduced, without written approval, from ACS Laboratory. The results of this report relate only to the material or product analyzed. Test results are confidential unless explicitly waived otherwise. ACS Laboratory is accredited to the ISO/IEC 17025:2017 Standard.



Certificate of Analysis

Compliance Test

Client Information:

Honest PP&D

1038 Arlington St

Orlando, Florida 32805

Order # HON250410-010004

Order Date: 2025-04-10

Sample # AAGP159

Batch # GYDHJWM252401

Batch Date: 2025-03-26

Extracted From: HEMP

Test Reg State: Georgia

Sampling Date: 2025-04-11

Lab Batch Date: 2025-04-11

Completion Date: 2025-04-16

Initial Gross Weight: 60.400 g



Heavy Metals - (GA) (Edible)

Specimen Weight: 252.050 mg

Passed
SOP13.048 (ICP-MS)

Dilution Factor: 198.373

Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)	Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)
Arsenic (As)	0.013	100	500	<LOQ	Lead (Pb)	0.007	100	500	<LOQ
Cadmium (Cd)	0.003	100	500	<LOQ	Mercury (Hg)	0.016	100	500	<LOQ



Residual Solvents (GA-4)

Specimen Weight: 17.200 mg

Passed
SOP13.039 (GCMS-HS)

Dilution Factor: 1.000

Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)	Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)
Butane	0.4167	53.6	800000	<LOQ	Heptane	0.0013	29.8	500000	<LOQ
Ethanol	0.0021	59.5	5000000	96366.940	Hexane	0.068	25	100000	<LOQ



Mycotoxins - (GA)

Specimen Weight: 586.540 mg

Passed
SOP13.007 (LCMS)

Dilution Factor: 2.560

Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)	Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)
Aflatoxin B1	3.0400E-1	1.9	20	<LOQ	Aflatoxin G2	2.7100E-1	1.9	20	<LOQ
Aflatoxin B2	7.7000E-2	1.9	20	<LOQ	Ochratoxin A	7.5400E-1	3.8	20	<LOQ
Aflatoxin G1	3.0400E-1	1.9	20	<LOQ					

Aixia Sun

Aixia Sun Lab Director/Principal Scientist

D.H.Sc., M.Sc., B.Sc., MT (AAB)

Definitions are found on page 1

This report shall not be reproduced, without written approval, from ACS Laboratory. The results of this report relate only to the material or product analyzed. Test results are confidential unless explicitly waived otherwise. ACS Laboratory is accredited to the ISO/IEC 17025:2017 Standard.





Certificate of Analysis

Compliance Test

Client Information:

Honest PP&D

**1038 Arlington St
Orlando, Florida 32805**

Order # HON250410-010004

Order Date: 2025-04-10

Sample # AAGP159

Batch # GYDHJWM252401

Batch Date: 2025-03-26

Extracted From: HEMP

Test Reg State: Georgia

Sampling Date: 2025-04-11

Lab Batch Date: 2025-04-11

Completion Date: 2025-04-16

Initial Gross Weight: 60.400 g



Pesticides - (GA)

Specimen Weight: 586.540 mg

Passed

SOP13.007 (LCMS)

Dilution Factor: 2.560

Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)	Analyte	LOD (ppb)	LOQ (ppb)	Action Level (ppb)	Result (ppb)
Abamectin	3.9900E-1	9.6	100	<LOQ	Fludioxonil	3.6000E-1	9.6	100	<LOQ
Acephate	1.4100E-1	9.6	100	<LOQ	Hexythiazox	1.1300E-1	9.6	100	<LOQ
Acequinocyl	2.1780E+0	9.6	100	<LOQ	Imazalil	2.5800E-1	9.6	100	<LOQ
Acetamiprid	1.4000E-1	9.6	100	<LOQ	Imidacloprid	4.0200E-1	9.6	100	<LOQ
Aldicarb	2.0300E-1	9.6	100	<LOQ	Kresoxim Methyl	1.8200E-1	9.6	100	<LOQ
Azoxystrobin	1.8800E-1	9.6	100	<LOQ	Malathion	2.2300E-1	9.6	100	<LOQ
Bifenazate	8.6000E-2	9.6	100	<LOQ	Metalaxyl	2.7000E-1	9.6	100	<LOQ
Bifenthrin	1.0000E-1	9.6	100	<LOQ	Methiocarb	1.1800E-1	9.6	100	<LOQ
Boscalid	5.9500E-1	9.6	100	<LOQ	Methomyl	6.4000E-2	9.6	100	<LOQ
Carbaryl	1.2200E-1	9.6	100	<LOQ	Mevinphos	9.3000E-2	9.6	100	<LOQ
Carbofuran	8.6000E-2	9.6	100	<LOQ	Myclobutanil	5.7300E-1	9.6	100	<LOQ
Chlorantraniliprole	8.4000E-2	9.6	100	<LOQ	Oxamyl	4.1000E-2	9.6	100	<LOQ
Chlordane	1.9000E+0	9.6	100	<LOQ	Paclobutrazol	1.8600E-1	9.6	100	<LOQ
Chlorpyrifos	1.0900E-1	9.6	100	<LOQ	Permethrin	6.2400E-1	9.6	100	<LOQ
Coumaphos	2.0600E-1	9.6	100	<LOQ	Phosmet	1.2700E-1	9.6	100	<LOQ
Cyfluthrin	9.8000E-1	9.6	100	<LOQ	Piperonylbutoxide	1.4900E-1	9.6	100	<LOQ
Cypermethrin	9.8500E-1	9.6	100	<LOQ	Prallethrin	1.4760E+0	9.6	100	<LOQ
Daminozide	1.6550E+0	9.6	100	<LOQ	Propiconazole	2.9400E-1	9.6	100	<LOQ
Diazinon	2.1200E-1	9.6	100	<LOQ	Propoxur	1.0000E-1	9.6	100	<LOQ
Dichlorvos	1.1300E+0	9.6	100	<LOQ	Pyridaben	1.4000E-1	9.6	100	<LOQ
Dimethoate	6.3000E-2	9.6	100	<LOQ	Spinetoram	4.2400E-1	9.6	100	<LOQ
Dimethomorph	2.5810E+0	9.6	100	<LOQ	Spiromesifen	1.2000E-1	9.6	100	<LOQ
Ethoprophos	1.5100E-1	9.6	100	<LOQ	Spirotetramat	2.1100E-1	9.6	100	<LOQ
Etofenprox	1.7200E-1	9.6	100	<LOQ	Spiroxamine	5.3300E-1	9.6	100	<LOQ
Etoxazole	8.6600E-1	9.6	100	<LOQ	Tebuconazole	2.3000E-1	9.6	100	<LOQ
Fenhexamid	5.8800E-1	9.6	100	<LOQ	Thiacloprid	1.7000E-1	9.6	100	<LOQ
Fenoxycarb	2.7400E-1	9.6	100	<LOQ	Thiamethoxam	1.7900E-1	9.6	100	<LOQ
Fipronil	3.1700E-1	9.6	100	<LOQ	Trifloxystrobin	1.3400E-1	9.6	100	<LOQ
Flonicamid	4.6600E-1	9.6	100	<LOQ					

Aixia Sun

Aixia Sun Lab Director/Principal Scientist

D.H.Sc., M.Sc., B.Sc., MT (AAB)

Definitions are found on page 1

This report shall not be reproduced, without written approval, from ACS Laboratory. The results of this report relate only to the material or product analyzed. Test results are confidential unless explicitly waived otherwise. ACS Laboratory is accredited to the ISO/IEC 17025:2017 Standard.



Zombi Death Drops

Sample ID: 2502AFL0223.1038

Internal Lab ID: ATL-26900

Client

Strain: Wicked Watermelon

Honest PPD

Matrix: Ingestible

Received: 02/10/2025

Lic. #

Type: Soft Chew

Completed: 03/03/2025

1038 Arlington St

Unit Weight: 5.2137 g

Batch#: GYDHJWM252401

Orlando, FL 32805



Summary

Test	Date Tested	Result
Batch		Pass
Cannabinoids	02/10/2025	Complete
Residual Solvents	02/20/2025	Pass
Microbials	02/24/2025	Pass
Pesticides	03/03/2025	Pass
Heavy Metals	03/06/2025	Pass
Foreign Matter	02/24/2025	Pass

Pass/Fail criteria based on CMTL sample testing Florida Department of Health Emergency Rule for 2020. Measures of uncertainty are considered when deciding PASS/FAIL results. Analysis Method: ATL-PHI-001

Basic Potency

Complete

0.16% Total THC	ND Total CBD	0.48% Total Cannabinoids
---------------------------	------------------------	------------------------------------

Analyte	LOQ	Result	Result	Analyte	LOQ	Result	Result
	%	%	mg/unit		%	%	mg/unit
CBDVa	0.01	ND	ND	9S-HHC	0.01	0.08	4.2
CBDV	0.01	ND	ND	9R-HHC	0.01	0.21	10.8
CBDa	0.01	ND	ND	Δ9-THCH	0.01	ND	ND
CBGa	0.01	ND	ND	Δ8-THCH	0.01	ND	ND
CBG	0.01	ND	ND	Δ9-THCP	0.01	0.03	1.8
CBD	0.01	ND	ND	Δ8-THCP	0.01	ND	ND
THCV	0.01	ND	ND	Δ8-THCO	0.01	ND	ND
Δ8-THCV	0.00	ND	ND	Δ9-THCO	0.01	ND	ND
THCVa	0.01	ND	ND	9(S)-HHCP	0.01	ND	ND
CBN	0.01	ND	ND	9(R)-HHCP	0.01	ND	ND
Δ9-THC	0.01	0.16	8.1	11-Hydroxy-THC	0.01	ND	ND
Δ8-THC	0.01	ND	ND	1(S)-H4-CBD	0.01	ND	ND
(6aR,9S)-d10-THC	0.01	ND	ND	1(R)-H4-CBD	0.03	ND	ND
(6aR,9R)-d10-THC	0.00	ND	ND	Δ9-THCB	0.00	ND	ND
CBC	0.01	ND	ND	Δ8-THCB	0.01	ND	ND
CBCa	0.01	ND	ND	CBDp	0.01	ND	ND
THCa	0.01	ND	ND	Total		0.48	24.9

Document ID: ATL-225, Rev 05, Effective Date: 03 October 2023; Date Tested: 02/10/2025

NT Not Tested Moisture Content	NT Not Tested Water Activity	Pass Foreign Matter
---	---	-------------------------------




Paulo Gonzalez
Laboratory Manager
03/03/2025

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com





Accurate Test Labs
1305 NW 65th Pl
Fort Lauderdale, FL 33309

(954) 515-0200
http://accuratetestlab.com
Lic#

QA Testing

2 of 4

Zombi Death Drops

Sample ID: 2502AFL0223.1038

Internal Lab ID: ATL-26900

Client

Strain: Wicked Watermelon

Honest PPD

Matrix: Ingestible

Received: 02/10/2025

Lic. #

Type: Soft Chew

Completed: 03/03/2025

1038 Arlington St

Unit Weight: 5.2137 g

Batch#: GYDHJWM252401

Orlando, FL 32805

Pesticides

Pass

Analyte	Limit	Result	Status	Analyte	Limit	Result	Status
	PPB	PPB			PPB	PPB	
Acephate	3000	ND	Pass	Fipronil	100.0	ND	Pass
Acequinocyl	2000	ND	Pass	Fludioxonil	3000	ND	Pass
Azoxystrobin	3000	ND	Pass	Imazalil	100.0	ND	Pass
Bifenazate	3000	ND	Pass	Kresoxim Methyl	1000	ND	Pass
Bifenthrin	500.0	ND	Pass	Malathion	2000	ND	Pass
Boscalid	3000	ND	Pass	Metalaxyl	3000	ND	Pass
Captan	3000	ND	Pass	Methiocarb	100.0	ND	Pass
Carbaryl	500.0	ND	Pass	Mevinphos	100.0	ND	Pass
Carbofuran	100.0	ND	Pass	Myclobutanil	3000	ND	Pass
Chlorantraniliprole	3000	ND	Pass	Naled	500.0	ND	Pass
Chlordane	100.0	ND	Pass	Paclobutrazol	100.0	ND	Pass
Chlorfenapyr	100.0	ND	Pass	Parathion Methyl	100.0	ND	Pass
Chlorpyrifos	100.0	ND	Pass	Pentachloronitrobenzene	200.0	ND	Pass
Clofentezine	500.0	ND	Pass	Permethrin	1000	ND	Pass
Coumaphos	100.0	ND	Pass	Phosmet	200.0	ND	Pass
Cyfluthrin	1000	ND	Pass	Piperonyl Butoxide	3000	ND	Pass
Cypermethrin	1000	ND	Pass	Propiconazole	1000	ND	Pass
Diazinon	200.0	ND	Pass	Propoxur	100.0	ND	Pass
Dichlorvos	100.0	ND	Pass	Pyrethrins	1000	ND	Pass
Dimethoate	100.0	ND	Pass	Pyridaben	3000	ND	Pass
Dimethomorph	3000	ND	Pass	Spiroxamine	100.0	ND	Pass
Ethoprophos	100.0	ND	Pass	Tebuconazole	1000	ND	Pass
Etofenprox	100.0	ND	Pass	Thiamethoxam	1000	ND	Pass
Etoxazole	1500.0	ND	Pass	Trifloxystrobin	3000	ND	Pass
Fenoxycarb	100.0	ND	Pass				

Date Tested: 03/03/2025

Action limits based on Florida Department of Agriculture and Consumer Services, Rule 5K-4.034 § 5(b)




Paulo Gonzalez
Laboratory Manager
03/03/2025

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com





Accurate Test Labs
1305 NW 65th Pl
Fort Lauderdale, FL 33309

(954) 515-0200
http://accuratetestlab.com
Lic#

QA Testing

3 of 4

Zombi Death Drops

Sample ID: 2502AFL0223.1038

Internal Lab ID: ATL-26900

Client

Strain: Wicked Watermelon

Honest PPD

Matrix: Ingestible

Received: 02/10/2025

Lic. #

Type: Soft Chew

Completed: 03/03/2025

1038 Arlington St

Unit Weight: 5.2137 g

Batch#: GYDHJWM252401

Orlando, FL 32805

Microbials

Pass

Analyte	LOD CFU/g	Result CFU/g	Status
Aerobic Bacteria	10	NR	NT
Aspergillus	1	Not Detected in 1g	Pass
Shiga toxin-producing E. Coli	1	Not Detected in 1g	Pass
Enterobacteriaceae	10	NR	NT
Salmonella	1	Not Detected in 1g	Pass
Staphylococcus aureus	10	NR	NT
Yeast & Mold	10	ND	Pass

Document ID: ATL-235, Rev 01, Effective Date: 13 May 2022

Date Tested: 02/24/2025

Pass/Fail criteria based on CMTL sample testing Florida Department of Health Emergency Rule for 2020. Measures of uncertainty are considered when deciding PASS/FAIL results. Analysis Method: ATL-PHI-002 and ATL-PHI-004.

Foreign Matter

Pass

Analyte	Limit % by weight	Result % by weight	Status
Packaging Contaminants	0.01	ND	Pass
Insects	0.01	ND	Pass
Feces	0.005	ND	Pass
Mold	0.01	ND	Pass
FM Other		ND	Pass
Hair	0.01	ND	Pass

Document ID:

Date Tested: 02/24/2025

Pass/Fail criteria based on CMTL sample testing Florida Department of Health Emergency Rule for 2020. Measures of uncertainty are considered when deciding PASS/FAIL results. Analysis Method: ATL-PHI-001

Heavy Metals

Pass

Analyte	LOQ µg/g	Limit µg/g	Result µg/g	Status
Arsenic	100	1500	ND	Pass
Cadmium	100	500	ND	Pass
Lead	100	500	ND	Pass
Mercury	100	3000	ND	Pass

Date Tested: 03/06/2025

Passed - Analyte/microbe is not detected or is at the level below the action limit per FL rule 64ER20-39, 5K-4.036, 5K-4.034. Failed - Analyte/microbe is at the level that equal or above the action limit per FL rule 64ER20-39, 5K-4.036, 5K-4.034



Paulo Gonzalez
Laboratory Manager
03/03/2025

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com





Accurate Test Labs
1305 NW 65th Pl
Fort Lauderdale, FL 33309

(954) 515-0200
<http://accuratetestlab.com>
Lic#

QA Testing

4 of 4

Zombi Death Drops

Sample ID: 2502AFL0223.1038

Internal Lab ID: ATL-26900

Client

Strain: Wicked Watermelon

Honest PPD

Matrix: Ingestible

Received: 02/10/2025

Lic. #

Type: Soft Chew

Completed: 03/03/2025

1038 Arlington St

Unit Weight: 5.2137 g

Batch#: GYDHJWM252401

Orlando, FL 32805

Residual Solvents

Pass

Analyte	LOQ	Limit	Result	Status
	µg/g	µg/g	µg/g	
1,1 Dichloroethene	0.7	8	ND	Pass
Ethylene Dichloride	0.2	2	ND	Pass
Acetone	59	750	ND	Pass
Acetonitrile	5	60	ND	Pass
Benzene	0.08	1	ND	Pass
Chloroform	0.2	2	ND	Pass
Ethanol	79	5000	ND	Pass
Ethyl Acetate	20	400	ND	Pass
Ethyl Ether	40	500	ND	Pass
Ethylene Oxide	0.4	5	ND	Pass
Heptane	40	5000	ND	Pass
Isopropanol	40	500	ND	Pass
Methanol	20	250	<20.000	Pass
Methylene Chloride	10	125	ND	Pass
Butane	71	5000	ND	Pass
Hexane	5	250	ND	Pass
Pentane	59	750	ND	Pass
Propane	165	5000	ND	Pass
Toluene	12	150	ND	Pass
Trichloroethene	2	25	ND	Pass
Xylenes	12	150	ND	Pass

ACCURATE
TEST LAB

Document ID: ATL-230, Rev 01, Effective Date: 13 May 2022

Date Tested: 02/20/2025

Action limits based on Florida Department of Agriculture and Consumer Services, Rule 5K-4.034 § 5(c); Expanded uncertainty = ±25% considered for PASS/FAIL; Test Method ATL-GCM-001



Paulo Gonzalez
Laboratory Manager
03/03/2025

Confident LIMS
All Rights Reserved
coa.support@confidentlims.com
(866) 506-5866
www.confidentlims.com

